gutflix
GI learning, personalised

Summary

DEEP LEARNING AND CAPSULE ENDOSCOPY: AUTOMATIC MULTI-BRAND PANENDOSCOPIC DETECTION OF VASCULAR LESIONS

Miguel Mascarenhas-Saraiva

UEG Week Copenhagen 2023 | Poster

Oesophagus

October 16, 2023

A multicentric retrospective study developed and validated a convolutional neural network model for automatic detection of vascular lesions across multiple capsule endoscopy device brands, achieving 95.0% accuracy on 35655 frames from 1022 exams.

KEY POINTS

- The CNN model demonstrated 86.4% sensitivity, 98.3% specificity, 95.2% positive predictive value, 95.0% negative predictive value, and 0.96 AUC-PR for detecting vascular lesions (angiectasia or varices) during capsule endoscopy.
- The model was trained and validated using frames from seven different capsule endoscopy device types (PillCam Colon 1, PillCam Crohn's, PillCam SB1, PillCam SB3, OMOM HD, Olympus, Mirocam), with all frames from individual patients assigned to the same dataset.
- The validation dataset contained 2433 normal mucosa frames and 933 vascular lesion frames, with the CNN processing 115 frames per second.
- The authors note this is the first proof-of-concept AI model for panendoscopic vascular lesion detection during capsule endoscopy, but acknowledge limitations including the retrospective design and potential selection bias, recommending prospective multicentric studies before clinical implementation.

VIEW THIS CONTENT ON GUTFLIX

https://gutflix.eu/search/d/b6ef267e-743c-11ee-8138-0242ac140004

AI-generated summary

This summary was generated by an AI large language model based on the content transcript. It is for informational purposes only and should not be considered a substitute for clinical judgment. Always rely on your professional expertise and the full clinical context when making clinical decisions.